



[www.insize.com](http://www.insize.com)



**7305-G3350  
DATA OUTPUT CABLE  
OPERATION MANUAL**



## Introduction

- 7305-G3350** data output cable  
 Black cable, length: 3m  
 Serial port signal  
 Interface type: Micro-USB




7305-G3350

## Communication protocol

- Baud: 9600bps  
 Data bit: 8  
 Stop bit: 1  
 Check bit: None  
 No flow control

- Data output format

1. Hex



2. ASCII

Data format example	
Display	Output
1.234	➡ +1.234
-1.234	➡ -1.234

## Connect Operation

After the data output cable is connected to the 3350 series digital micrometer, connect to the computer



## Operating Instructions

### Drive installation:

- Driver download address:  
[www.insize.com/download/PL2303\\_Prolific\\_DriverInstaller\\_v110.zip](http://www.insize.com/download/PL2303_Prolific_DriverInstaller_v110.zip)
- Unzip package, double-click the installation program, and Install

### Switching transmission mode:

- Open any serial port send-receive software (such as serial port assistant), select the correct communication protocol(9600 8 N 1),and open the serial port
- Single transmit instruction: RD  
Clear instruction: 99 (hexadecimal characters)
- Can also press the DATA button of digital indicator to transmit data